

SIEMENS

Siemens Healthcare Diagnostics, a global leader in clinical diagnostics, provides healthcare professionals in hospital, reference, and physician office laboratories and point-of-care settings with the vital information required to accurately diagnose, treat, and monitor patients. Our innovative portfolio of performance-driven solutions and personalized customer care combine to streamline workflow, enhance operational efficiency, and support improved patient outcomes.

Actin, BCS, BEP, Berichrom, Ci-Trol, Dade, Data-Fi, Enzygnost, Hepzyme, INNOVANCE, Innovin, Multifibren, Pathromtin, PFA, ProC, Thromboclotin, Thromborel, Turbiquant, vWF Ag, and all associated marks are trademarks of Siemens Healthcare Diagnostics Inc. Sysmex is a trademark of Sysmex Corporation. All other trademarks and brands are the property of their respective owners.

Product availability may vary from country to country and is subject to varying regulatory requirements. Please contact your local representative for availability.

This brochure is intended for use outside U.S. only (O.U.S.). Display or promotion to U.S. audience is prohibited.

Order No. A91DX-110902-XC1-4A00  
11-2011 | All rights reserved  
© 2011 Siemens Healthcare Diagnostics Inc.



[www.siemens.com/diagnostics](http://www.siemens.com/diagnostics)

# Hemostasis Reagents Portfolio

Comprehensive reagent options to power up your lab

**Global Siemens Headquarters**  
Siemens AG  
Wittelsbacherplatz 2  
80333 Muenchen  
Germany

**Global Siemens Healthcare Headquarters**  
Siemens AG  
Healthcare Sector  
Henkestrasse 127  
91052 Erlangen  
Telephone: +49 9131 84-0  
Germany  
[www.siemens.com/healthcare](http://www.siemens.com/healthcare)

**Global Division**  
Siemens Healthcare Diagnostics Inc.  
511 Benedict Avenue  
Tarrytown, NY 10591-5005  
USA  
[www.siemens.com/diagnostics](http://www.siemens.com/diagnostics)

[www.siemens.com/diagnostics](http://www.siemens.com/diagnostics)

Answers for life.

# Hemostasis Reagents Portfolio

For more than 30 years, we have held the distinction of being recognized as the global leader in hemostasis testing. Our comprehensive portfolio of instruments and reagent offerings enables custom-fit solutions to laboratories of all sizes. Our assays offer a broad selection of various testing solutions to support physicians in making sound diagnostic and therapeutic decisions. The hemostasis assay portfolio ranges from standard PT and APTT testing to the breakthrough von Willebrand Factor activity testing technology found in our INNOVANCE® VWF Ac Assay. No matter how routine or specialized your testing may be, we have the solutions to ensure quality testing, standardize results, and offer solutions to meet the needs of your laboratory.

			Instrument Availability							Comments	
			Systems and Analyzers			Sysmex® Systems					
Reagent Name	Reagent Description		Catalog No.	Package Size	BCS® BCS® XP	BCT™	BFT II	CA-7000	CA-1500		CA-560
PT	Thromborel® S Reagent	Thromborel S Reagent is prepared from human placental tissue factor combined with calcium chloride and stabilizers. The reagent contains minimal residual clotting factors, such as prothrombin or factors VII or X, for clear definition of factor deficiencies and steep factor assay curves. Because of its high sensitivity to these coagulation factors, the reagent is suitable for monitoring oral anticoagulant therapy.  Thromborel S Reagent exhibits good correlation with the WHO international reference thromboplastin preparation. With the Thromborel S Reagent and the appropriate deficient plasma, it is possible to determine activity of coagulation factors II, V, VII, and X. The reagent differentiates abnormal plasmas, even in the mildly pathological range.	OUHP29 OUHP49	10 x 4 mL 10 x 10 mL	●	●	●	●	●	●	●
	Dade® Innovin® Reagent	Dade Innovin Reagent is prepared from purified recombinant human tissue factor produced in <i>E. coli</i> , combined with synthetic phospholipids, calcium, buffers, and stabilizers. It is highly sensitive to extrinsic factor deficiencies and oral anticoagulant-treated patient plasma samples. The sensitivity of Dade Innovin Reagent is very similar to that of the WHO human brain reference thromboplastin. It is insensitive to therapeutic levels of heparin, which, in combination with high sensitivity to coagulation factors, makes Dade Innovin Reagent ideal for monitoring oral anticoagulant therapy and differentiating abnormal plasmas, even in the mildly pathological range.	B4212-40 B4212-50 B4212-100	10 x 4 mL 10 x 10 mL 12 x 20 mL	●	●	●	●	●	●	●
APTT	Dade Actin® Activated Cephaloplastin Reagent	Dade Actin Activated Cephaloplastin Reagent has moderate sensitivity to factor deficiencies (VIII, IX, XI, and XII) in the intrinsic system. It is the ideal choice for institutions requiring a moderate screening APTT reagent for routine testing. Dade Actin has low heparin sensitivity, allowing the monitoring of heparin therapy even with high heparin dosage. Dade Actin Reagent has moderate sensitivity to lupus anticoagulants.	B4218-1 B4218-2	10 x 2 mL 10 x 10 mL	●	●	●	●	●	●	●
	Dade Actin FS Activated PTT Reagent	Dade Actin FS Activated PTT Reagent is a highly sensitive reagent for the detection of factor deficiencies (VIII, IX, XI, XII) of the intrinsic system. With moderate sensitivity to lupus anticoagulants and high sensitivity to heparin, it fulfills all requirements of routine coagulation testing.	B4218-20 B4218-100	10 x 2 mL 10 x 10 mL	●	●	●	●	●	●	●
	Dade Actin FSL Activated PTT Reagent	Dade Actin FSL Activated PTT Reagent exhibits increased sensitivity to lupus anticoagulants and moderate heparin sensitivity. The reagent shows good factor sensitivity to detect clinically significant deficiencies of the intrinsic system.	B4219-1 B4219-2	10 x 2 mL 10 x 10 mL	●	●	●	●	●	●	●
	Pathromtin® SL Reagent	Pathromtin SL Reagent exhibits high sensitivity to lupus anticoagulants, factor deficiencies, and heparin.	OQGS29 OQGS35	10 x 5 mL 20 x 5 mL	●	●	●	●	●	●	●
Fibrinogen	Multifibren® U Reagent	Multifibren U Reagent is a bovine thrombin reagent used in the modified Clauss determination of fibrinogen for the detection of hereditary or acquired hypo- and hyperfibrinogenemia and dysfibrinogenemia. The reagent is insensitive to heparin up to 2.0 U/mL and has a wide measuring range of 0.80–12.00 g/L.	OWZG19 OWZG23	10 x 2 mL 10 x 5 mL	●	●	●	●	●	●	●
	Dade Thrombin Reagent	Dade Thrombin Reagent is an effective reagent for use in the determination (Clauss method) of fibrinogen in the detection of hereditary or acquired hypo- and hyperfibrinogenemia, dysfibrinogenemia, and afibrinogenemia. The reagent offers long stability after reconstitution.	B4233-25 B4233-27	10 x 1 mL 10 x 5 mL				●	●	●	●
	Dade Fibrinogen Determination Reagent	The Dade Fibrinogen Determination Reagent consists of Dade Thrombin Reagent, Fibrinogen Standard, and Dade Owren's Veronal Buffer for use in the determination of fibrinogen (Clauss method) in the detection of hereditary or acquired hypo- and hyperfibrinogenemia, dysfibrinogenemia, and afibrinogenemia. The reagent offers long stability after reconstitution.	B4233-15SY	Kit				●	●	●	●
Thrombin Time/ Batroxobin Time	BC Thrombin Reagent	BC Thrombin Reagent is utilized for the determination of the thrombin time. The BC Thrombin Reagent is suitable for monitoring of fibrinolytic therapy, screening for disorders of fibrin formation, in suspected cases of severe fibrinogen deficiency states, and for differentiation between heparin-induced prolongation of the thrombin time and disorders of fibrinogen formation.  Thrombin time is found to be prolonged not only due to disorders in fibrin polymerization, but also due to the presence of heparin. Differentiation can be achieved using Batroxobin Reagent.	OWNA11	Kit	●	●					
	Thromboclotin® Reagent	Thromboclotin Reagent is intended for the determination of thrombin time in citrated human plasma. The reagent is suitable for monitoring of fibrinolytic therapy, screening for disorders of fibrin formation, in suspected cases of severe fibrinogen deficiency states, and for differentiation between heparin-induced prolongation of the thrombin time and disorders of fibrinogen formation.  Thrombin time is found to be prolonged not only due to disorders in fibrin polymerization, but also due to the presence of heparin. Differentiation can be achieved using Batroxobin Reagent.	281007	10 x 10 mL	●	●	●		●	●	●

	Reagent Name	Reagent Description
Thrombin Time/ Batroxobin Time	Test Thrombin Reagent	Test Thrombin Reagent is intended for the determination of thrombin time in citrated human plasma. The reagent is suitable for monitoring of fibrinolytic therapy, screening for disorders of fibrin formation, in suspected cases of severe fibrinogen deficiency states, and for differentiation between heparin-induced prolongation of the thrombin time and disorders of fibrinogen formation.  Thrombin time is found to be prolonged not only due to disorders in fibrin polymerization, but also due to the presence of heparin. Differentiation can be achieved using Batroxobin Reagent.
	Batroxobin Reagent	Batroxobin Reagent is a snake venom-based reagent intended for the determination of the batroxobin time. It is ideal for monitoring fibrinolytic therapy by determination of fibrinogen/fibrin degradation products, diagnosis of afibrinogenemia and dysfibrinogenemia, and elucidation of prolonged thrombin times in cases of suspected presence of heparin.
Single Factors	Coagulation Factor II Deficient Plasma	Coagulation Factor II Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor II (prothrombin). It is manufactured by immunoabsorption and contains a residual factor concentration of <1% prothrombin and normal levels of fibrinogen and other extrinsic clotting factors. Coagulation Factor II Deficient Plasma was designed to be used in combination with Dade Innovin or Thromborel S Reagents.
	Coagulation Factor V Deficient Plasma	Coagulation Factor V Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor V. It is manufactured by immunoabsorption and contains a residual factor concentration of <1% factor V and normal levels of fibrinogen and other extrinsic clotting factors. Coagulation Factor V Deficient Plasma was designed to be used in combination with Dade Innovin or Thromborel S Reagents.
	Coagulation Factor VII Deficient Plasma	Coagulation Factor VII Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor VII. It is manufactured by immunoabsorption and contains a residual factor concentration of <1% factor VII and normal levels of fibrinogen and other extrinsic clotting factors. Coagulation Factor VII Deficient Plasma was designed to be used in combination with Dade Innovin or Thromborel S Reagents.
	Coagulation Factor VIII Deficient Plasma	Coagulation Factor VIII Deficient Plasma is a lyophilized reagent immunoabsorbed from human plasma. It contains < 1% residual factor VIII and normal levels of fibrinogen and other intrinsic coagulation factors. It is used for the determination of factor VIII activity in the diagnosis of congenital and acquired deficiencies of factor VIII.
	Coagulation Factor IX Deficient Plasma	Coagulation Factor IX Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor IX (hemophilia B). With a residual factor concentration of <1%, the reagent is ideal for the monitoring of substitution therapy. Coagulation Factor IX Deficient Plasma was designed to be used in combination with Dade Actin, Dade Actin FS, Dade Actin FSL, or Pathromtin SL Reagents.
	Coagulation Factor X Deficient Plasma	Coagulation Factor X Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor X. It is manufactured by immunoabsorption and contains a residual factor concentration of <1% factor X and normal levels of fibrinogen and other extrinsic clotting factors. Coagulation Factor X Deficient Plasma was designed to be used in combination with Dade Innovin or Thromborel S Reagents.
	Coagulation Factor XI Deficient Plasma	Coagulation Factor XI Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor XI. The reagent has a residual factor concentration of <1% factor XI and was designed to be used in combination with Dade Actin, Dade Actin FS, Dade Actin FSL, or Pathromtin SL Reagents.
	Coagulation Factor XII Deficient Plasma	Coagulation Factor XII Deficient Plasma is a human plasma-based reagent for the detection of hereditary or acquired deficiencies of factor XII. The reagent has a residual factor concentration of <1% factor XII and was designed to be used in combination with Dade Actin, Dade Actin FS, Dade Actin FSL, or Pathromtin SL Reagents.
	Berichrom Factor XIII Kit	The Berichrom Factor XIII Kit is a chromogenic, quantitative assay for the detection of hereditary or acquired factor XIII deficiencies. The chromogenic activity reagent is also utilized for the monitoring of patients undergoing factor XIII substitution therapy.
	Factor VIII Chromogenic Assay	The Factor VIII Chromogenic Assay is recommended for factor FVIII determination in therapeutic factor FVIII preparations and the detection of hereditary or acquired factor VIII deficiencies. The chromogenic method is insensitive to heparin at levels of <10 IU/mL.
von Willebrand Factors	INNOVANCE® VWF Ac Kit	The INNOVANCE VWF Kit is a sensitive, reliable, and convenient test system for direct determination of VWF activity. It employs an advanced new technology, allowing the assay to mimic the way in which VWF binds to glycoprotein Ib (GPIb) the reaction, the major VWF receptor protein on platelets. Latex particles are coated with an antibody against GPIb, to which recombinant GPIb is added. The addition of patient plasma induces a VWF-dependent agglutination, which is detected turbidimetrically. Because the recombinant receptor protein includes two gain-of-function mutations, the assay does not require ristocetin.
	BC von Willebrand Reagent	BC von Willebrand Reagent provides a simple, rapid, and automated procedure for the determination of the ristocetin cofactor activity of von Willebrand factor. The reagent, which provides a rapid measurement time of 250 seconds, is sensitive to types 1, 2, and 3 of von Willebrand disease (except VWD 2N) and is the recommended screening method for von Willebrand disease.
	von Willebrand Reagent	von Willebrand Reagent is a manual, quantitative activity method sensitive to types 1, 2, and 3 of von Willebrand disease (except VWD 2N). The ristocetin cofactor assay is recommended for the screening of von Willebrand disease.
	vWF Ag® Kit	vWF Ag Kit contains is a quantitative, automated immunoassay used to determine the differentiation of quantitative versus qualitative von Willebrand factor deficiencies. It is sensitive to types 1 and 3 VWF deficiencies and offers a wide measuring range of 2–600%.

Catalog No.	Package Size	Instrument Availability							Comments
		Systems and Analyzers			Sysmex Systems				
		BCS BCS XP	BCT	BFT II	CA-7000	CA-1500	CA-560	CS-2000i CS-2100i	
OWHM13	10 x 5 mL			●	●	●	●	●	
OOUV21	2 x 5 mL	●	●	●	●	●	●	●	
OSGR13	3 x 1 mL	●	●	●	●	●		●	
ORSM19	8 x 1 mL	●	●	●	●	●		●	
OTXV13	3 x 1 mL	●	●	●	●	●	●	●	
OTXW17	8 x 1 mL	●	●	●	●	●	●	●	
OTXX17	8 x 1 mL	●	●	●	●	●		●	
OTXY13	3 x 1 mL	●	●	●	●	●		●	
OSDF13	3 x 1 mL	●	●	●	●	●		●	
OSDG13	3 x 1 mL	●	●	●	●	●		●	
OWSU11	Kit	●	●					●	
B4238-40	Kit	●	●		●	●		●	
OPHL03	Kit	●			●	●	●	●	
OUBD37	5 x 4 mL	●	●					●	
OUBD23	5 x 2 mL								manual method
OPAB03	Kit	●	●		●	●	●	●	

	Reagent Name	Reagent Description
Thrombophilia	LA 1 Screening Reagent	LA 1 Screening Reagent contains dilute Russell's viper venom and low phospholipids for use in the simplified DRVVT as a screening test for lupus anticoagulants. The LA 1 Screening Reagent was designed to be used in conjunction with the LA 2 Confirmation Reagent.
	LA 2 Confirmation Reagent	LA 2 Confirmation Reagent is a simplified dilute Russell's viper venom test rich in phospholipids, making it ideal for the confirmation of lupus anticoagulants. The LA 2 Confirmation Reagent was designed to be used in conjunction with the LA 1 Screening Reagent.
	ProC® Global Kit	ProC Global Kit is a coagulometric screening reagent for the protein C pathway. It provides a determination of the anticoagulatory capacity of the protein C system. The heparin-insensitive reagent is useful in screening individuals affected by thrombophilia. ProC Global Kit is sensitive to deficiencies of factor V Leiden and proteins C and S, certain lupus anticoagulants, and high factor VIII levels.
	ProC Ac R Kit	The ProC Ac R Kit, a dilute Russell's viper venom test with a sensitivity and specificity of >99%, screens for APC resistance due to the presence of factor V Leiden in patient samples. The reagent is insensitive to heparin and is not influenced by high levels of factor VIII.
	Protein C Reagent	Protein C Reagent is a coagulometric reagent used for the quantitative determination of protein C activity. The reagent is suitable for the detection of hereditary or acquired protein C deficiencies.
	Protein S Ac Reagent	Protein S Ac Reagent, a coagulometric activity reagent, is utilized for the detection of hereditary or acquired protein S deficiencies.
	Berichrom® Protein C Kit	The Berichrom Protein C Kit, a chromogenic activity assay, is utilized for the detection of hereditary or acquired protein C deficiency types. The assay is also used for the monitoring of substitution therapy with protein C concentrates in congenital protein C deficiency. The Berichrom Protein C Kit is less susceptible to interfering substances than a clotting assay.
	INNOVANCE Antithrombin Kit	The INNOVANCE Antithrombin Kit is an automated chromogenic assay for the quantitative determination of functional antithrombin. The human factor Xa-based reagent has minimal interference with heparin cofactor II and thrombin inhibitors such as hirudin. The ready-to-use, liquid reagents provide excellent precision and reliability.
	Berichrom Antithrombin III (A) Kit	The Berichrom Antithrombin III (A) Kit is a chromogenic activity assay for the detection of hereditary or acquired antithrombin deficiency and monitoring of patients undergoing substitution therapy. The heparin co-factor-independent lyophilized reagent uses bovine thrombin and exhibits no interference with anti-FXa anticoagulants (e.g., rivaroxaban).
Heparin	Berichrom Heparin Kit	The Berichrom Heparin Kit, a chromogenic, factor Xa-based activity assay, is utilized for the monitoring of heparin therapy and the determination of unfractionated (UF) and low molecular weight (LMW) heparin in patient samples. The reagent is not influenced by platelet factor IV complexing of heparin and offers a detection limit of 0.05 IU/mL in UF heparin.
Fibrinolysis	Berichrom α2-Antiplasmin Kit	The Berichrom α2-Antiplasmin Kit is utilized for the determination of α2-antiplasmin and the detection of hereditary or acquired α2-antiplasmin deficiencies.
	Berichrom Plasminogen Kit	Berichrom Plasminogen Kit, a chromogenic activity test system, is used for the determination of plasminogen and the detection of hereditary or acquired plasminogen deficiencies.
	Berichrom PAI Kit	The Berichrom PAI Kit is a chromogenic test system for the determination of PAI (Plasminogen Activator Inhibitor) levels as an indicator of a thrombophilic state and hypofibrinolysis. The reagent is not influenced by α2-antiplasmin or FDP.
D-Dimer	INNOVANCE D-Dimer Kit	The INNOVANCE D-Dimer Kit is a rapid, highly precise, and sensitive test system for the determination of D-dimer. It offers high diagnostic sensitivity of >98% for exclusion of VTE (venous thromboembolism). With its extended assay range, D-dimer levels can be used for the diagnosis and monitoring of patients with disseminated intravascular coagulopathy (DIC), as well as for the monitoring of anticoagulation treatment and pregnancy-related coagulopathies (e.g., pre-eclampsia and HELLP syndrome).
	Dade Dimertest Latex Assay	The Dade Dimertest Latex Assay is a rapid agglutination test system using latex particles coated with a specific D-dimer monoclonal antibody. Dimertest is intended for the qualitative or semi-quantitative evaluation of cross-linked fibrin degradation products containing D-dimers.
	D-Dimer Latex Beads	D-Dimer Latex Beads are latex particles coated with a specific D-dimer monoclonal antibody used in the qualitative or semi-quantitative evaluation of cross-linked fibrin degradation products containing D-dimers.

		Instrument Availability							Comments
		Systems and Analyzers			Sysmex Systems				
Catalog No.	Package Size	BCS BCS XP	BCT	BFT II	CA-7000	CA-1500	CA-560	CS-2000i CS-2100i	
OQGP17	10 x 2 mL	•	•	•	•	•	•	•	
OQGR13	10 x 1 mL	•	•	•	•	•	•	•	
OQLS13	Kit	•	•	•	•	•		•	
OPBC03	Kit	•			•	•		•	
OQYG11	Kit	•	•	•	•	•	•	•	
OPAP03	Kit	•	•		•	•		•	
OUVV15 OUVV17	Kit	•	•		•	•	•	•	
OPFH03 OPFH05	Small Kit Large Kit	•			•	•	•	•	
OWWR17 OWWR15	Small Kit Large Kit	•	•		•	•	•	•	
OWLD11	Kit	•	•		•	•	•	•	
OUBU15	Kit	•	•		•	•		•	
OUCA17	Kit	•	•		•	•		•	
OWOA15	Kit	•							
OPBP03 OPBP07	150-Tests Kit 300-Tests Kit	•	•		•	•	•	•	
B4233-60	Kit								manual method
B4233-61	1 x 2 mL								manual method

	Reagent Name	Reagent Description
Other	Enzygnost® TAT micro Kit	The Enzygnost TAT micro is an ELISA assay for thrombin-antithrombin complex determination. The reagent is utilized for the diagnosis of hypercoagulability (e.g., in DIC).
	Enzygnost F1+2 (monoclonal) Kit	The Enzygnost F1+2 (monoclonal) is an ELISA assay for prothrombin fragment 1 and 2 determination. The reagent is utilized for the diagnosis of hyper- and hypocoagulable states on BEP Systems.
	Complement Reagents	The Complement Reagent Kit is a functional-activity assay for the determination of total complement. It is useful in the diagnosis of hereditary and acquired defects of the complement system and for monitoring response to treatment.
	Berichrom C1-Inhibitor Kit	The Berichrom C1-Inhibitor Kit, a human C1 esterase-based assay, determines the presence of C1 inhibitors in patient samples. The reagent offers a fast turnaround time-to-result of <10 minutes and detects hereditary or acquired deficiencies of the C1 inhibitor (e.g., in angioneurotic edema). This chromogenic activity reagent is used for the diagnosis of angioneurotic edema and for monitoring substitution or steroid therapy in angioneurotic edema.
	INNOVANCE ETP Kit	The INNOVANCE ETP Kit is a global hemostasis-function test system to assess endogenous thrombin potential (ETP). Several parameters are commonly used to describe ETP, from which the area under the curve (AUC) and peak height (Cmax) have been shown to be of diagnostic relevance: <ul style="list-style-type: none"> <li>Increased AUC has been demonstrated to correlate with an increased risk for recurrent venous thrombosis after discontinuation of anticoagulation.</li> <li>Increased AUC and Cmax have been observed due to prothrombin mutation G20210A.</li> <li>AUC and Cmax are known to be decreased under anticoagulant treatment with vitamin K antagonists.</li> <li>Reduction of AUC and Cmax have been demonstrated during treatment with argatroban (direct thrombin inhibitor).</li> </ul>
Supplementary	Calcium Chloride Solution	Calcium Chloride Solution is used as a supplementary reagent for various coagulation tests.
	Dade Hepzyme® Reagent	Dade Hepzyme Reagent is used as a heparin neutralizer in plasma to rule out heparin contamination in coagulation testing.
	Owren's Veronal Buffer	Owren's Veronal Buffer is a dilution buffer for coagulation testing.
	INNOVANCE D-Dimer Diluent	INNOVANCE D-Dimer Diluent is a liquid used for dilution of samples with elevated D-dimer concentrations when running the INNOVANCE D-Dimer Assay.
	Imidazole Buffer Solution	Imidazole Buffer Solution is used as a supplementary reagent for various coagulation assays run on the BCT and BFT II Systems.
	Kaolin Suspension	Kaolin Suspension is used as a supplementary reagent for various assays for the BFT II System.
Platelets	INNOVANCE PFA P2Y Cartridges	INNOVANCE PFA P2Y is an easy and reliable test for the detection of P2Y12 receptor blockade in patients undergoing therapy with a P2Y12 receptor blockade antagonist.
	Dade PFA Collagen/EPI Test Cartridges	The Dade PFA Collagen/EPI Test Cartridge is used for the detection of platelet dysfunction; screening for intrinsic platelet defects, von Willebrand disease, or exposure to platelet inhibiting agents; presurgical screening for bleeding risk; and monitoring of aspirin effect and DDAVP. It is sensitive to all types of von Willebrand disease (except 2N), hereditary platelet defects, low platelet count (<150,000/μL), and to aspirin and anti-GP IIb/IIIa antagonists.
	Dade PFA Collagen/ADP Test Cartridges	The Dade PFA Collagen/ADP Test Cartridges are utilized for the differentiation of aspirin effect on platelets versus other platelet dysfunctions. It is insensitive to aspirin, yet sensitive to VWD, low platelet counts, and other platelet dysfunctions.
	Dade PFA Trigger Solution	Dade PFA Trigger Solution is an isotonic buffer solution used for triggering the membrane for cartridges for the PFA Systems.
	Dade Cluster Platelet Aggregation Reagents	Dade Cluster Platelet Aggregation Reagents—consisting of collagen, ADP, and epinephrine—are utilized in platelet aggregation studies for screening of inherited and acquired platelet dysfunction.
Standards and Calibrators	Standard Human Plasma	Standard Human Plasma is citrated normal human pooled plasma intended for the calibration of various coagulation and fibrinolysis assays. Standard human plasma is calibrated against the respective WHO standard, where available.
	PT-Multi Calibrator	The PT-Multi Calibrator comprises a set of six plasmas intended for the direct calibration of prothrombin time (PT) in INR and % of norm. The calibrators are also suitable for the determination of a local ISI value. The single plasma levels have calibrated values for Innovin and Thromborel S Reagents on each individual instrument.
	Fibrinogen Calibrator Kit	The Fibrinogen Calibrator Kit comprises a set of six plasmas used to prepare reference curves for the fibrinogen assay by the modified Clauss method using Siemens Healthcare Diagnostics' Multifibren U Reagent. (Fibrinogen levels 1–6 have a range of approximately 0.6–9.0 g/L.)
	Berichrom Heparin UF Calibrator	The Berichrom Heparin UF Calibrator is for use in the preparation of an unfractionated heparin calibration curve with the Berichrom Heparin Kit. It is calibrated against the fifth WHO standard for unfractionated heparin (human pooled plasma containing UFH ≤1.3 U/mL).
	Berichrom Heparin LMW Calibrator	The Berichrom Heparin LMW Calibrator is for use in preparation of a LMW heparin calibration curve with the Berichrom Heparin Kit. It is calibrated against the second WHO standard for LMWH (human pooled plasma containing LMWH ≤1.5 U/mL).

Catalog No.	Package Size	Instrument Availability							Comments
		Systems and Analyzers			Sysmex Systems				
		BCS BCS XP	BCT	BFT II	CA-7000	CA-1500	CA-560	CS-2000i/CS-2100i	
OWMG15	Kit								ELISA/BEP® Systems
OPBD03	Kit								ELISA/BEP Systems
OWZD11	Kit	●	●						
OUIA15	Kit	●	●		●			●	
OPGA	2 x 50 Tests	●							
ORHO37	10 x 15 mL	●	●	●	●	●	●	●	
B4240-10	10 x 1 mL	●	●	●	●	●	●	●	
B4234-25	10 x 15 mL	●	●		●	●	●	●	
OPBR03	10 x 5 mL	●	●		●	●	●	●	
OQAA33	6 x 15 mL		●	●					
OQAB42	1 x 50 mL			●					
B4170-22	1 x 20 Cartridges								PFA-100® System, INNOVANCE PFA-200 System
B4170-21A	1 x 20 Cartridges								PFA-100 System, INNOVANCE PFA-200 System
B4170-21A	1 x 20 Cartridges								PFA-100 System, INNOVANCE PFA-200 System
B4170-50	3 x 11 mL								PFA-100 System, INNOVANCE PFA-200 System
B4236-1	Kit		●						manual test
ORKL17	10 x 1 mL	●	●	●	●	●	●	●	
OPAT03	6 x 1 mL	●	●	●	●	●	●	●	
OQVK11	6 x 1 mL	●	●	●	●	●	●	●	
OPCC03	6 x 1 mL	●	●		●	●	●	●	
OPCA03	6 x 1 mL	●	●		●	●	●	●	

	Reagent Name	Reagent Description
Controls	Control Plasma N	Control Plasma N is citrated normal human pooled plasma from selected healthy blood donors. Control Plasma N is used for the assessment of the precision and analytical deviation of various analytes in the normal range. This control provides assigned values for the respective available analytes.
	Control Plasma P	Control Plasma P is citrated human plasma from selected healthy blood donors. Control Plasma P is a precision and accuracy control intended to monitor the performance of various parameters in the pathological range. The control provides assigned values for the respective available analytes.
	Dade Ci-Trol® 1, 2, and 3 Controls	Dade Ci-Trol Level 1, 2, and 3 Controls are intended for use as precision and accuracy controls in the normal, mid, and upper therapeutic ranges for the global assays. The controls provide assigned values for the respective available analytes.
	Dade Ci-Trol Coagulation Control Level 1, 2, and 3	Dade Ci-Trol Level 1, 2, and 3 Controls are composed of citrated human pooled plasma from selected healthy blood donors. They are intended for use as controls in the normal, mid, and upper therapeutic ranges.
	V.E.Q. Coag A Control	Global Control consist of citrated normal human pooled plasma from selected healthy blood donors. The control is utilized for precision control of coagulation tests in the normal range.
	V.E.Q. Coag B Control	Global Control consist of citrated human pooled plasma from selected healthy blood donors. The control is utilized for precision control of coagulation tests in the pathological range.
	Dade Data-Fi® Abnormal Fibrinogen Control Plasma	Dade Data-Fi Abnormal Fibrinogen Control Plasma is a control derived from human plasma. It is used to assess accuracy and precision of Dade Fibrinogen Determination Reagents in the low range.
	LA Control Low Control	LA Control Low is a low-positive control for lupus anticoagulant clotting assays using LA 1 Screening and LA 2 Confirmation Reagents.
	LA Control High Control	LA Control High is a high-positive control for lupus anticoagulant clotting assays using LA 1 Screening and LA 2 Confirmation Reagents.
	ProC Control Plasma	ProC Control Plasma is an assayed intralaboratory control to estimate precision and analytical deviation of the ProC line of tests in the pathological range.
	Ci-Trol Heparin Control Low	Ci-Trol Heparin Control Low is a low-level control using the activated partial thromboplastin time (APTT).
	Ci-Trol Heparin Control High	Ci-Trol Heparin Control High is a high-level control using the activated partial thromboplastin time (APTT).
	INNOVANCE D-Dimer Controls	INNOVANCE D-Dimer Controls 1 and 2 are assayed controls for the assessment of precision and analytical bias in the normal and pathological range for the determination of D-dimer with the INNOVANCE D-Dimer Assay.
	Berichrom Heparin UF Control 1	Berichrom Heparin UF Control 1 is a precision and accuracy control used to monitor the performance of the Berichrom Heparin Assay in the therapeutical unfractionated heparin range.
	Berichrom Heparin UF Control 2	Berichrom Heparin UF Control 2 is a precision and accuracy control used to monitor the performance of the Berichrom Heparin Assay in the subtherapeutical unfractionated heparin range.
	Berichrom Heparin LMW Control 1	Berichrom Heparin LMW Control is a precision and accuracy control used to monitor the performance of the Berichrom Heparin Assay in the therapeutical low-molecular-weight heparin range.
Berichrom Heparin LMW Control 2	Berichrom Heparin LMW Control 2 is a precision and accuracy control used to monitor the performance of the Berichrom Heparin Assay in the subtherapeutical low-molecular-weight heparin range.	

Catalog No.	Package Size	Instrument Availability							Comments
		Systems and Analyzers			Sysmex Systems				
		BCS BCS XP	BCT	BFT II	CA-7000	CA-1500	CA-560	CS-2000i CS-2100i	
ORKE41	10 x 1 mL	•	•	•	•	•	•	•	
OUPZ17	10 x 1 mL	•	•	•	•	•	•	•	
291070 291071 291072	10 x 1 mL 10 x 1 mL 10 x 1 mL	•	•	•	•	•	•	•	
B4244-10 B4244-20 B4244-30	20 x 1 mL 20 x 1 mL 20 x 1 mL	•	•	•	•	•	•	•	
291043	10 x 1 mL	•	•	•	•	•	•	•	
291044	10 x 1 mL	•	•	•	•	•	•	•	
B4233-22	10 x 1 mL				•	•	•	•	
OQWE11	6 x 1 mL	•	•	•	•	•	•	•	
OQWD11	6 x 1 mL	•	•	•	•	•	•	•	
OQKE17	6 x 1 mL	•	•	•	•	•		•	
B4224-50	10 x 1 mL	•	•		•	•	•		
B4224-60	10 x 1 mL	•	•		•	•	•		
OPDY03	2 x 5 x for 1 mL	•	•		•	•	•	•	
OPBY03	6 x 1 mL	•	•		•	•	•	•	
OPBZ03	6 x 1 mL	•	•		•	•	•	•	
OPCD03	6 x 1 mL	•	•		•	•	•	•	
OPCB03	6 x 1 mL	•	•		•	•	•	•	